



MicroQuant™ *Pseudomonas* *paraeruginosa*, low CFU (Pack of 5)

9027-LQ-PACK™

Description

MicroQuant™

MicroQuant™ *Pseudomonas paraeruginosa* (low CFU) is a quantitated, best-in-class control developed to support USP General Chapter <61>. This ISO 17034 reference material is provided as a pack containing 5 vials of cryopreserved (100 to 1,000 CFU per pellet, supporting 10 assays per pellet) and 5 vials of rehydration buffer. This convenient, single-use format rehydrates immediately and is easy to use and store.

Strain designation: R. Hugh 813

Deposited As: *Pseudomonas aeruginosa* (Schroeter) Migula

Type strain: Yes

Shipping information: 5 vials of cryopreserved pellets and 5 vials of rehydration buffer

Storage Conditions

Storage conditions: 2°C to 8°C

Intended Use

This product is intended for laboratory research use only. It is not intended for any animal or human therapeutic use, any human or animal consumption, or any diagnostic use.

Reference material produced under an ISO 17034 accredited process.

MicroQuant™ *Pseudomonas paraeruginosa*, low CFU (Pack of 5)

9027-LQ-PACK
BSL 2

Product Sheet

ATCC determines the biosafety level of a material based on our risk assessment as guided by the current edition of *Biosafety in Microbiological and Biomedical Laboratories (BMBL)*, U.S. Department of Health and Human Services. It is your responsibility to understand the hazards associated with the material per your organization's policies and procedures as well as any other applicable regulations as enforced by your local or national agencies.

ATCC highly recommends that appropriate personal protective equipment is always used when handling vials. For cultures that require storage in liquid nitrogen, it is important to note that some vials may leak when submersed in liquid nitrogen and will slowly fill with liquid nitrogen. Upon thawing, the conversion of the liquid nitrogen back to its gas phase may result in the vial exploding or blowing off its cap with dangerous force creating flying debris. Unless necessary, ATCC recommends that these cultures be stored in the vapor phase of liquid nitrogen rather than submersed in liquid nitrogen.

Certificate of Analysis

For batch-specific test results, refer to the applicable certificate of analysis that can be found at www.atcc.org.

Handling Procedures

1. Remove the desired number of cryopreserved reference culture vials (gray crimp) and rehydration buffer vials (green crimp) from 2-8°C storage and place into a biosafety cabinet. One vial of rehydration buffer should be used for each vial of reference culture.

MicroQuant™ *Pseudomonas paraeruginosa*, low CFU (Pack of 5)

Product Sheet

9027-LQ-PACK

2. Uncap each vial. Add 1.0 mL of rehydration buffer to the reference culture vial. Recap the vial and allow the pellet to dissolve for 30 seconds.
3. Vortex on high for 30 seconds to ensure the suspension is well mixed. The suspension is now useable and contains 100-1000 CFU per 1.0 mL. The suspension can be used immediately or stored at 2-8°C for up to 8 hours.
4. Uncap the vial and pipette 100 µL of the resuspended reference culture directly on a non-selective media agar plate. If following a specific pharmacopeial chapter, use the appropriate media specified.
Note: ATCC quantitation results were obtained using 100 mm Tryptic Soy Agar plates with 1.5% Tryptone, 0.5% Soytone, 0.5% Sodium Chloride, and 1.5% Agar (plates were procured from Teknova®).
5. Use a sterile plate spreader (not provided) to spread the sample evenly over the surface of the media plate. Do not use glass beads. Incubate the plates in an incubator at the appropriate temperature (see the product page or product sheet). If following a specific pharmacopeial chapter, use the incubation conditions specified.
6. Following incubation, count the number of visible colonies on the plate.

Notes

For applications outside of compendial assays, please refer to the product page for ATCC 9027 for standard growth conditions.

Material Citation

If use of this material results in a scientific publication, please cite the material in the following manner: MicroQuant™ *Pseudomonas paraeruginosa*, low CFU (Pack of 5) (ATCC 9027-LQ-PACK)

References

References and other information relating to this material are available at www.atcc.org.

MicroQuant™ *Pseudomonas paraeruginosa*, low CFU (Pack of 5)

9027-LQ-PACK

Warranty

The product is provided 'AS IS' and the viability of ATCC® products is warranted for 30 days from the date of shipment, provided that the customer has stored and handled the product according to the information included on the product information sheet, website, and Certificate of Analysis. For living cultures, ATCC lists the media formulation and reagents that have been found to be effective for the product. While other unspecified media and reagents may also produce satisfactory results, a change in the ATCC and/or depositor-recommended protocols may affect the recovery, growth, and/or function of the product. If an alternative medium formulation or reagent is used, the ATCC warranty for viability is no longer valid. Except as expressly set forth herein, no other warranties of any kind are provided, express or implied, including, but not limited to, any implied warranties of merchantability, fitness for a particular purpose, manufacture according to cGMP standards, typicality, safety, accuracy, and/or noninfringement.

Disclaimers

This product is intended for laboratory research use only. It is not intended for any animal or human therapeutic use, any human or animal consumption, or any diagnostic use. Any proposed commercial use is prohibited without a [license from ATCC](#).

While ATCC uses reasonable efforts to include accurate and up-to-date information on this product sheet, ATCC makes no warranties or representations as to its accuracy. Citations from scientific literature and patents are provided for informational purposes only. ATCC does not warrant that such information has been confirmed to be accurate or complete and the customer bears the sole responsibility of confirming the accuracy and completeness of any such information.

This product is sent on the condition that the customer is responsible for and assumes all risk and responsibility in connection with the receipt, handling, storage, disposal, and use of the ATCC product including without limitation taking all appropriate safety and handling precautions to minimize health or environmental risk. As a condition of receiving the material, the customer agrees that any activity

MicroQuant™ *Pseudomonas paraeruginosa*, low CFU (Pack of 5)

9027-LQ-PACK

Product Sheet

undertaken with the ATCC product and any progeny or modifications will be conducted in compliance with all applicable laws, regulations, and guidelines. This product is provided 'AS IS' with no representations or warranties whatsoever except as expressly set forth herein and in no event shall ATCC, its parents, subsidiaries, directors, officers, agents, employees, assigns, successors, and affiliates be liable for indirect, special, incidental, or consequential damages of any kind in connection with or arising out of the customer's use of the product. While reasonable effort is made to ensure authenticity and reliability of materials on deposit, ATCC is not liable for damages arising from the misidentification or misrepresentation of such materials.

Please see the material transfer agreement (MTA) for further details regarding the use of this product. The MTA is available at www.atcc.org.

Copyright and Trademark Information

© ATCC 2023. All rights reserved.

ATCC is a registered trademark of the American Type Culture Collection.

Revision

This information on this document was last updated on 2025-03-31

Contact Information

ATCC

10801 University Boulevard

Manassas, VA 20110-2209

USA

US telephone: 800-638-6597

Worldwide telephone: +1-703-365-2700

Email: tech@atcc.org or contact your local distributor