ATCC Medium:	4 Rabbit Blood	l Agar
---------------------	----------------	--------

Blood Agar Base (BD 211037).....40 g DI Water.....950 ml

After mixing blood agar base and DI water, autoclave at 121° C and cool to 48-50°C.

Aseptically add:

Rabbit Blood......50 ml

Gently mix and dispense as required.